

Draft Practice Standard: Electronic Mechanical Pump Selection for Home Administration of Infusion Medications

Introduction

The practice standards of NHIA represent a consensus of professional judgment, expert opinion, and documented evidence. They provide guidance and direction to NHIA members and other audiences who affect the home infusion industry, and the patients served. Their use may help to comply with federal and state laws and regulations, meet accreditation requirements, and improve patient care. They are written to establish reasonable goals, to be progressive and challenging, yet attainable as best practices in applicable home or alternate site settings. They should not be viewed as NHIA requirements. The use of NHIA's practice standards by members and other practitioners should be assessed and adapted based on independent judgment.

Purpose

The purpose of infusion device selection is to ensure patient safety and minimize adverse events, avoid unplanned hospitalization and emergency room visits, and prevent disruptions in treatment. The Pharmacist and Nurse are responsible for participating in the selection of an appropriate method of administration to be used in delivering parenteral medications to patients infusing medications at home. NHIA has developed a practice standard to help clinicians assess situations and determine when an electronic mechanical pump is recommended to administer an infusion medication.

Practice Standard Development Procedure

An area of practice was identified where information and direction are necessary to promote general standards and quality. The topic of appropriate selection of infusion device has generated a need among practitioners for authoritative advice. There was sufficient experience upon which to base a practice standard and it was relevant to the practice of a significant portion of NHIA's members. The NHIA Quality and Standards Committee, comprised of home infusion professionals, performed a comprehensive review of available evidence. Committee members who are experienced clinical and research experts collaborated on the topic, reviewing industry trends and other data sources such as membership, community, best practices, industry research, review of published literature, or other practice guidelines. A draft practice standard was developed and reviewed by the committee members and revisions to the draft were incorporated based on consensus.

Target Audience

The target audience includes clinicians, regulatory agencies, reimbursement professionals, and industry stakeholders.

Practice Standard for Electronic Mechanical Pump Selection

Situations where the use of an electronic mechanical pump is strongly recommended to infuse medications in the home:

- Continuously infused medications with a narrow therapeutic index requiring a strictly controlled infusion rate to avoid toxicity and achieve the desired response
 - Example situation: Continuous infusions of inotropic medications
- Continuous or extended infusions (≥ 3 hours) where dose titration or rate adjustments are required to achieve a clinical response or to avoid an adverse drug reaction.
 - Example situation: Parenteral nutrition formulas with titration/ramping parameters
 - Example situation: Narcotic/opioid infusions with boluses prescribed
- Continuous or extended subcutaneous medication infusions.

Situations where a clinician should consider the use of an electronic mechanical pump for medication administration:

- Administration of medications prescribed on a dosing frequency that promotes adherence to the prescribed dosing schedule.
 - Example situation: An antibiotic administered intermittently 3 or more times per day and dispensed as a multi-dose container for use over 24-48 hours.
- Infusion of medications where the administration period and/or self-administration are not otherwise easily achieved.
 - Example situation: An extended infusion of a beta-lactam antibiotic over 3 or more hours
- Administration of infusion solutions that require an infusion rate > 250 mL/hr over > 2 hours or > 500 mL
 - Example situation: Hydration solution with electrolytes administered at 500 mL/hr x 1000 mL
- Administration through a 'small bore catheter'
 - Example situation: 3 Fr or smaller

Comments/Contact: NHIA welcomes feedback and suggestions for content to incorporate into future revisions and editions. Contact: standards@nhia.org.

DISCLAIMER: The National Home Infusion Association (NHIA) produces educational resources to aid good clinical practice that reflects the input of its members and experienced clinicians in the field. The information offered in NHIA resources is intended as a guide for information purposes only and does not replace or remove clinical judgment or the professional care and duty necessary for each specific situation. While great effort has been made to assure all information is complete and accurate as of the time this resource was issued, given the continuously evolving health care environment and the particular circumstances of individual cases, no assurance can be given that the information is entirely complete or accurate in every conceivable respect (and, as such, NHIA and its board members, committee/work group members, officers and employees disclaim all liability for the accuracy or completeness of this resource, and disclaim all warranties, express or implied to its incorrect use).